

---

# TWL Calibration Data Summary

---

**Serial Number:** 0000958

**Full-Scale Flow Rate:** 80  
(yields 20mA output)

**Zero Flow Rate:** (yields 4mA output)

**Bridge Resistance:** 350.9 ohms

**Full-Scale Bridge Output:** 2.0841 mV/V

**Calibration Resistor (R7):** 47337 ohms

**Calibration Voltage:** 0.8859 volts

**Calibration Flow Rate:** 75.29

**Calibration Current:** 19.06 mA

**Zero Switch Setting:** 0

**TB3 Jumper Position:** + (POSITIVE)

**Span Switch Setting - Undamped:** 0

**Span Switch Setting - Damped:** 8

**Additional Information:**

---